



Medline Canada, Corp.
3-8690 Escarpment Way
Milton, Ontario L9T 0M1

1.905.636.2100
1.800.396.6996

medline.com

Product Information

Glove Name Aloetouch® 3G Synthetic Gloves
Description Powder-Free Stretch Vinyl Synthetic Coated Examination Gloves
Non-Sterile, Ambidextrous with Film Opening

US Office: Medline Industries, Inc.
Northfield, IL, USA
1-800-MEDLINE
www.medline.com

Canada Office: Medline Canada Corp.
8690 Escarpment Way, Unit #3
Milton, ON L9T 0M1
1-800-396-6996

Product Numbers		Packaging
MDS195173	X-Small	100 Each/Box
MDS195174	Small	10 Boxes/Case
MDS195175	Medium	1000 Each/Case
MDS195176	Large	(90 ea/bx, 10 bx/cs for XL)
MDS195177	X-Large	

Product Not Made with Natural Rubber Latex

Physical Properties and Quality Characteristics

Characteristics:	Glove Color/Finish	Green/Smooth
	Glove Length	240 mm
	Finger Thickness	0.17mm (6.7 mils)
	Palm Thickness	0.14mm (5.5 mils)
	Primary Material	Polyvinyl Chloride (PVC) with UHMW PVC Resin
	Accelerators	None
	Cuff	Beaded
Pinholes:	Acceptable Quality Level (AQL)	2.5
	ASTM Standard D5250	2.5
Tensile Strength:	Min	15 MPa
	ASTM Standard D5250	11 MPa
Ultimate Elongation:	Min	540%
	ASTM Standard D5250	300%
Viral Penetration Test:	ASTM Standard F1671	Complies with F1671 Test Method
Natural Rubber Latex Protein	ASTM Standard D5712	Not applicable; Latex-Free
Content:		
Residual Powder Content:	ASTM Standard D6124	Less than 2 mg/glove

Aloetouch and Medline are registered trademarks of Medline Industries, Inc.

© Medline Industries, Inc. not to be reproduced or distributed without prior written permission of Medline Industries, Inc.

Updated 01/2017

Original on file