



Product Specifications

Stellar Supreme Nitrile Examination Gloves



Measurements

Material	Cuff thickness	Palm Thickness	Finger thickness	Length	Donning	Colour	Cuff	Texture
Nitrile	min 0.06	min 0.07	min 0.08	240 mm	Powder Free	Blue	Beaded	Fingertip Textured

Item #	Size	Glove Weight (g)	Palm Width	Case Weight (lbs)		Before Aging	After Aging
002-E10206-XS	X Small	2.5 +/- 0.3	70 +/- 10	8.4	Tensile Strength (Mpa)	Min 15	Min 14
002-E10206-S	Small	3.0 +/- 0.3	80 +/- 10	8.9	Elongation (%)	Min 500	Min 400
002-E10206-M	Medium	3.5 +/- 0.3	95 +/- 10	9.5			
002-E10206-L	Large	4.0 +/- 0.3	110 +/- 10	10.6			
002-E10206-XL	Extra Large	4.5 +/- 0.3	115 +/- 10	11.7			

Classification

Particular	Chemical	Level	EN ISO 374-4:2019 Degradation (%)
EN ISO 374-1 : 2016+A1:2018/Type B	40% Sodium Hydroxide (K)	6	1.4
EN ISO 374-1 : 2016+A1:2018/Type B	30% Hydrogen Peroxide (P)	6	24.0
EN ISO 374-1 : 2016+A1:2018/Type B	37% Formaldehyde (T)	6	10.9

EN ISO 374-5 : 2016	Protection Against Bacteria and Fungi	Pass
EN ISO 374-5 : 2016	Protection Against Viruses	Pass

Storage: To be stored in a dry environment, avoid direct sunlight

Shelf Life: 5 year from date of manufacturing

Packaging: 200 Gloves / Box x 10 Boxes / Case= 2,000 Gloves / Case

Country of Origin: China

Acceptances: Comply with CFIA and FDA guidelines

Health Canada licenced Class II medical device